




MINOS Status Report

Steve Hahn



- This week
 - Live time: 80.3%
 - POT fraction: 84.2%
- Power supply and temperature problems:
 - Saturday 1 FE power supply (Wiener) tripped, reset remotely
 - Sunday 3 FE power supplies tripped, reset worked for a short period but tripped again
 - Monday traced back to power supply temp limit of 92° F. Now raised to 94°. Has been raised from 90° to 92° to 94°.
 - Temps have been rising in NUMI underground for months. New chilled water system now supposed to start Jan. 3. But air cooling seems to have problems, also; would be good to have this checked in addition.
 - Thanks to Benton, Donnatella, Howard Budd



MINOS Status Report

Steve Hahn



- This week
 - Live time: 96.2%
 - POT fraction: 95.4%

